

0106N008		WBC	RBC	HB	HCT	MCV	MCH	MCHC	PLT	RDW	PDW
MINDRAY BC (5150&5800&5000&6800)	MEAN	4.7	4.6	14.0	43.0	94.4	30.8	32.6	217		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	3.0	5.7	2.5	2.0	30		
	MEAN-2SD	4.2	4.3	13.2	40.0	88.7	28.3	30.7	187		
	MEAN+2SD	5.2	4.8	14.9	46.0	100.1	33.3	34.6	248		
MINDRAY BC (2600-2800-3000-3600)	MEAN	4.8	4.4	14.0	40.1	90.8	31.8	35.0	187		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.8	5.4	2.5	2.1	26		
	MEAN-2SD	4.3	4.2	13.2	37.3	85.4	29.2	32.9	161		
	MEAN+2SD	5.2	4.7	14.9	42.9	96.3	34.3	37.1	213		
MYTHIC ORPHEE	MEAN	4.9	4.2	14.4	40.7	96.9	34.1	35.2	165		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.9	2.9	5.8	2.7	2.1	23		
	MEAN-2SD	4.4	4.0	13.5	37.9	91.1	31.4	33.1	142		
	MEAN+2SD	5.4	4.5	15.2	43.6	102.7	36.9	37.4	188		
NIHON KOHDEN(G&ALFA)	MEAN	5.2	4.2	14.6	41.4	99.6	35.0	35.1	161		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.2	0.9	2.9	6.0	2.8	2.1	23		
	MEAN-2SD	4.7	3.9	13.7	38.5	93.7	32.2	33.0	138		
	MEAN+2SD	5.8	4.4	15.4	44.3	105.6	37.8	37.3	183		
PENTRA	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
PROKAN	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
QUINTOS	MEAN	4.9	4.5	14.1	42.0	93.1	31.3	33.6	211		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	30		
	MEAN-2SD	4.4	4.2	13.3	39.1	87.5	28.8	31.6	181		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.7	33.8	35.6	241		
SAMSUNG	MEAN	4.9	4.5	14.1	42.0	93.1	31.3	33.6	211		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	30		
	MEAN-2SD	4.4	4.2	13.3	39.1	87.5	28.8	31.6	181		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.7	33.8	35.6	241		
SIEMENS ADVIA	MEAN	4.5	4.5	14.5	39.0	86.5	32.1	37.1	153		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.9	2.7	5.2	2.6	2.2	21		
	MEAN-2SD	4.1	4.2	13.6	36.2	81.3	29.5	34.9	132		
	MEAN+2SD	5.0	4.8	15.3	41.7	91.6	34.6	39.3	174		
SWEALAB	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
SYSMEX (KX21& Kx21N- K800-K1000- K4500)	MEAN	4.9	4.7	14.2	42.1	89.7	30.2	33.7	214		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.9	2.9	5.4	2.4	2.0	30		
	MEAN-2SD	4.4	4.4	13.3	39.1	84.3	27.8	31.7	184		
	MEAN+2SD	5.4	5.0	15.0	45.0	95.1	32.6	35.7	244		
SYSMEX (X-XP-XS-XT-XN SERIES)	MEAN	5.0	4.6	14.1	41.8	91.7	30.8	33.6	196		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.5	2.5	2.0	27		
	MEAN-2SD	4.5	4.3	13.2	38.9	86.2	28.3	31.6	168		
	MEAN+2SD	5.5	4.8	14.9	44.8	97.2	33.3	35.6	223		
MARRIAGE CENTERS	MEAN	5.0	4.7	13.4	41.3	87.9	28.6	32.5	206		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.3	2.3	1.9	29		
	MEAN-2SD	4.5	4.4	12.6	38.4	82.7	26.3	30.5	177		
	MEAN+2SD	5.5	5.0	14.2	44.2	93.2	30.8	34.4	234		

0106N008		WBC	RBC	HB	HCT	MCV	MCH	MCHC	PLT	RDW	PDW
BECKMAN CULTER	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
BIOTA	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
CELLDYN	MEAN	4.2	4.4	14.3	43.2	98.6	32.7	33.1	208		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.4	0.3	0.9	3.0	5.9	2.6	2.0	29		
	MEAN-2SD	3.8	4.1	13.4	40.1	92.7	30.1	31.2	178		
	MEAN+2SD	4.7	4.6	15.2	46.2	104.5	35.3	35.1	237		
DIAGON	MEAN	4.7	4.3	13.6	39.5	90.8	31.4	34.5	225		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.8	5.4	2.5	2.1	32		
	MEAN-2SD	4.2	4.1	12.8	36.7	85.3	28.9	32.5	194		
	MEAN+2SD	5.2	4.6	14.5	42.2	96.2	33.9	36.6	257		
DIATRON	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
DIRUI	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
FORTRESS	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
HEMRAJ& RHYTO	MEAN	5.1	4.1	12.8	39.0	96.1	31.5	32.8	185		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.2	0.8	2.7	5.8	2.5	2.0	26		
	MEAN-2SD	4.6	3.8	12.0	36.3	90.3	29.0	30.9	159		
	MEAN+2SD	5.6	4.3	13.6	41.7	101.8	34.0	34.8	211		
HIPERION	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
HYCELL	MEAN	4.9	4.4	14.2	41.7	94.9	32.3	34.1	211		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.9	2.9	5.7	2.6	2.0	30		
	MEAN-2SD	4.4	4.1	13.3	38.8	89.2	29.8	32.0	182		
	MEAN+2SD	5.4	4.7	15.1	44.6	100.6	34.9	36.1	241		
MEDONIC	MEAN	5.3	4.5	13.7	41.7	92.3	30.4	32.9	215		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.5	2.4	2.0	30		
	MEAN-2SD	4.8	4.2	12.9	38.7	86.8	27.9	30.9	185		
	MEAN+2SD	5.8	4.8	14.5	44.6	97.8	32.8	34.9	245		
MELET	MEAN	4.9	4.5	14.1	42.0	92.7	31.1	33.6	200		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	2.9	5.6	2.5	2.0	28		
	MEAN-2SD	4.4	4.3	13.3	39.1	87.2	28.6	31.6	172		
	MEAN+2SD	5.4	4.8	14.9	44.9	98.3	33.6	35.6	228		
MICROSS	MEAN	4.9	4.6	14.0	42.6	93.6	30.8	32.9	158		
	%CV	5.0	6.0	3.0	3.5	3.0	4.0	3.0	7.0		
	2SD	0.5	0.3	0.8	3.0	5.6	2.5	2.0	22		
	MEAN-2SD	4.4	4.3	13.2	39.6	88.0	28.3	30.9	136		
	MEAN+2SD	5.4	4.8	14.8	45.6	99.2	33.2	34.8	180		